

|   |         |
|---|---------|
| MODE OPERATOIRE / OPERATING PROCEDURE                                     | VERSION |
| <b>DETECTION OF MEMBRANE PERMEABILITY BY<br/>7-AAD (LIVE DEAD MARKER)</b> | A       |

### Reagents and materials

- 96-well plates with conical bottoms.
- phosphate buffered saline (PBS) 1x.
- 7-AAD stock solution (1mg/ml in 1x PBS) (cat. No 420403).
- Trypsin 1x.
- PBS without calcium or magnesium.
- Paraformaldehyde at 4%.
- Actinomycin D (AD) stock solution (1mg/ml in 1x PBS).

Note: provide a culture condition inducing a positive cell death control.

### Staining of cells in suspension

This test is carried out on  $1 \times 10^5$ - $2 \times 10^5$  cells per sample.

- Centrifuge the samples at 300 G for 5 min. Remove the supernatant.
- Resuspend the pellets in 200  $\mu$ l of 1x PBS and transfer in a 96-well plate.
- Centrifuge at 300 G for 5 min. Remove the supernatant.
- Add 100  $\mu$ l of 1x PBS containing 20  $\mu$ l/ml of 7-AAD.
- Incubate 20 min at 4°C in the dark.
- Centrifuge at 300 G for 5 min. Remove the supernatant.
- Wash in 200  $\mu$ l of 1x PBS containing 20  $\mu$ g/ml of AD.
- If immediate acquisition of the samples is possible, resuspend the cells in 200  $\mu$ l of AD solution at 20  $\mu$ g/ml in 1x PBS, transfer them to FACS tubes and acquire on the cytometer.
- Otherwise, fix the samples in 200  $\mu$ l of 1x PBS containing 20  $\mu$ g/ml of AD and 1% PFA, 15 min at 4°C in the dark. Then preferably acquire samples within 24 hours.

### Labeling of adherent cells

To determine the percentage of apoptosis, it is necessary to analyze the adherent cells and the cells present in the supernatant.

The labeling can be carried out directly in the culture medium.

- Add the volume of 7-AAD at 1mg/ml necessary to obtain a final concentration of the marker in the culture of 20  $\mu$ g/ml. Incubate for 20 min at 4°C in the dark.
- Collect the supernatant from each culture and place it in a specifically annotated tube.
- Centrifuge the cells of the supernatant at 300 G for 5 min. Remove the supernatant.
- Wash the wells with 200  $\mu$ l of 1x PBS without calcium or magnesium.
- Detach the adherent cells with trypsin containing 20  $\mu$ g/ml of AD.
- Transfer the adherent cells into the tube containing the cells of the corresponding supernatant.
- Centrifuge the cells at 300 G for 5 min. Remove the supernatant.
- Wash the pellets in 200  $\mu$ l of 1x PBS containing 20  $\mu$ g/ml of AD.
- If immediate acquisition of samples is possible, resuspend the cells in 200  $\mu$ l of 1x PBS containing 20  $\mu$ g/ml of AD, and acquire on the cytometer
- Otherwise, fix the samples in 200  $\mu$ l of PBS containing 20  $\mu$ g/ml of ACet 1% PFA. Acquire within 24 hours.

|   |         |
|---|---------|
| MODE OPERATOIRE / OPERATING PROCEDURE                                     | VERSION |
| <b>DETECTION OF MEMBRANE PERMEABILITY BY<br/>7-AAD (LIVE DEAD MARKER)</b> | A       |

**Technical data sheet 7-AAD Viability Dye – Beckman Coulter**  
A07704 150 tests; 3 mL 20 µL/test - For In Vitro Diagnostic Use

### INTENDED USE

7-AAD Viability Dye is a chemical coloring agent in solution, which is designed for mono- or multiparametric analysis using flow cytometry. It permits non-viable cells to be identified and to be excluded from (viable) cells of interest in the analysis of human leucocytes.

### PRINCIPLE

This test depends upon the ability of 7-Amino-Actinomycin D (7-AAD) to insert itself between the tops of successive Cytosine (C) / Guanine (G) bases of the DNA double strand, when the interior of the cell and the nuclear chromatin are accessible.

Staining of the DNA double strand is undertaken by incubating the sample with the 7-AAD Viability Dye. The red cells are then removed by lysis and the leucocytes are analyzed by flow cytometry.

The flow cytometer analyzes light diffusion and the fluorescence of cells. It makes possible the localization of cells within the electronic window defined on a histogram, which correlates the orthogonal diffusion of light (Side Scatter or SS) and the fluorescence detected using a 620 or 675 nm “band pass” filter, corresponding to 7-AAD staining. Other histograms combining two of the different parameters available on the cytometer are also used in the gating stage.

The fluorescence of cells thus gated is analyzed in order to exclude the positively stained events (non-viable) from the unstained ones (viable). The results can be expressed as a percentage of non-fluorescent events in relation to all the events taken into account by the gating.

### WARNING AND PRECAUTIONS

1. Do not use the reagent beyond the expiry date.
2. Do not freeze.
3. Let it come to room temperature (18–25°C) before use.
4. Minimize exposure to light.
5. Avoid microbial contamination of the reagents, or false results may occur.
6. This ready for use reagent contains 0.005% (weight/volume) of 7-AAD.  
In pure form, 7-AAD is potentially carcinogenic. Though extremely diluted in the present formulation, this constituent may retain all or part of its harmful effects. It is therefore advisable to avoid all contact with the skin, mucosa and eyes.  
Use this reagent following the precautions for use (notably: the wearing of gloves, gown and protective glasses).
7. This ready-for-use reagent contains 1% (volume/volume) of Dimethyl Sulfoxide (DMSO).  
In pure form, DMSO is an irritant. Though extremely diluted in the present formulation, this constituent may retain all or part of its harmful effects. It is therefore advisable to avoid all contact with the skin, mucosa and eyes.  
Use this reagent following the precautions for use (notably: the wearing of gloves, gown and protective glasses).
8. Never pipette by mouth.
9. 7-AAD is a chromophore subject to degradation by exposure to light. It is packaged in a vial which is opaque to the wavelengths to which it is sensitive, so protecting it during storage.  
Avoid continuous exposure to light during the incubation stages and reduce light exposure of the samples once staining has been undertaken.



|   |         |
|---|---------|
| MODE OPERATOIRE / OPERATING PROCEDURE                                     | VERSION |
| <b>DETECTION OF MEMBRANE PERMEABILITY BY<br/>7-AAD (LIVE DEAD MARKER)</b> | A       |

10. All blood samples must be considered as potentially infectious and must be handled with care (in particular: the wearing of protective gloves, gowns and goggles). Never pipette by mouth.
11. Blood tubes and disposable material used for handling should be disposed of in ad hoc containers intended for incineration.

### GHS HAZARD CLASSIFICATION

7-AAD Viability Dye WARNING

H316 Causes mild skin irritation.

P332+P313 If skin irritation occurs: Get medical advice/attention.

Dimethyl Sulfoxide 0.5 - 1.5% Safety Data Sheet is available at [techdocs.beckmancoulter.com](http://techdocs.beckmancoulter.com)

### STORAGE AND STABILITY

The 7-AAD Viability Dye must be kept at between 2 and 8°C and protected from light, before and after the vial has been opened.

Stability of closed vial: see expiry date on vial.

Stability of opened vial: the reagent is stable for 60 days.

### SAMPLES

Venous blood must be taken using sterile tubes containing an EDTA salt as the anticoagulant.

The samples should be kept at room temperature (18–25°C) and not shaken. The sample should be homogenized by gentle agitation prior to taking the test sample.

The samples must be analyzed within 24 hours of venipuncture.

### PREPARATION OF SAMPLES

The concentration of leucocytes in the sample must be less than necessary, dilute in PBS to bring the leucocyte concentration to

The procedure uses 100 µL of a sample, tube prediluted or otherwise. NB:

- Adjustment of the intensity of fluorescence corresponding to 7-AAD can be undertaken using the staining of a sample of whole fresh blood and of non-viable stabilized cells in the same tube (see image in appendix). Under these conditions, the voltage corresponding to the detection of 7-AAD must be adjusted such that viable events (i.e., whole fresh blood) appear on the first decade of an SS histogram, depending on the 7-AAD (see example in the appendix).
- Do not use reagents containing permeating or fixating agents during the procedure in order to prevent artifactually positive staining.

### EVIDENCE OF DETERIORATION

Any change in the physical appearance of the reagents may indicate deterioration and the reagent should not be used.

In case of packaging deterioration or if data obtained show some performance alteration, please contact your local distributor or use the following e-mail address :

[immuno-techsup@beckmancoulter.com](mailto:immuno-techsup@beckmancoulter.com)

### MATERIALS NEEDED BUT NOT SUPPLIED WITH KIT:

- Sampling tubes and material necessary for sampling.
- Automatic pipettes with disposable tips for 20, 100 and 500 µL.
- Plastic haemolysis tubes.



|   |         |
|---|---------|
| MODE OPERATOIRE / OPERATING PROCEDURE                                     | VERSION |
| <b>DETECTION OF MEMBRANE PERMEABILITY BY<br/>7-AAD (LIVE DEAD MARKER)</b> | A       |

- Red cell lysis reagent. For example: IOTest3 Lysing Solution (Ref. A07799).
- Buffer (PBS: 0.01M sodium phosphate; 0.145 M sodium chloride; pH 7.2).
- Centrifuge.
- Automatic agitator (Vortex type).
- Flow cytometer.

## PROCEDURE

1. Add 20µL of the 7-AAD Viability Dye solution to each test tube.  
B59275-AC 3 of 72
2. Add 100µL of the test sample (i.e., the equivalent of approximately  $5 \times 10^5$  cells). Vortex the tubes gently.
3. Incubate for 15 to 20 minutes at room temperature (18–25°C), protected from light.
4. Then perform lysis of the red cells, if necessary, by following the recommendations of the lysis reagent used.  
By way of example, if one wishes to use the IOTest3 Lysing Solution (Ref. A07799) add 2 mL of the working solution (1X), vortex immediately and incubate for 10 minutes at room temperature, protected from light.  
If the sample does not contain red cells, do not undertake the lysis stage, but add 0.5 to 1 mL of PBS.
5. Preparations should be analyzed within 1 hour.

Note: In all cases, keep the preparations between 2 and 8°C and protected from light.

## EXPECTED RESULTS

Each laboratory must compile a list of reference values based upon a group of healthy donors from the local population. This must be done by taking age, sex and ethnic group into account, as well as any other potential regional differences.

In our laboratories, the less-than-6-hour whole blood samples of 10 healthy adults were treated using the reagent described above. The results obtained for the numeration of the events of interest with this reagent (dead leucocytes, ie 7-AAD positive) are given in the table below:

| Positive Target Leucocytes | Number | Mean (%) | SD   | CV (%) |
|----------------------------|--------|----------|------|--------|
| 7-AAD <sup>+</sup>         | 10     | 7.67     | 7.54 | 98     |

## SPECIFICITY

Actinomycins are the active biological constituents of a chromophore (2-Amino-4,6Dimethylphenoxazone-3) and of cyclic pentapeptides (3). They are antibiotics of bacterial origin characterized historically in soil Ascomycetes. The actinomycins form stable complexes with double strand deoxyribonucleic acid (DNA), but do not form this type of complex either with double strand ribonucleic acid (RNA) or with RNA-DNA hybrids, or with single strand DNA or RNA.

7-Amino-Actinomycin D (7-AAD) is an analogue of Actinomycin containing an amine group substituted in position 7 of the chromophore. 7-AAD inserts itself between the tops of the Cytosine/Guanine bases of DNA (4).

The spectral properties of 7-AAD make it a compound which is particularly well suited to flow cytometry analysis (3). The maximum absorption of the 7-AAD/DNA complex is compatible with the blue excitation wavelength of 488 nm for cytometers fitted with an argon laser (5). The peak fluorescence emission in the deep red band (635 to 675 nm) of the 7-AAD/DNA complex permits



|   |         |
|---|---------|
| MODE OPERATOIRE / OPERATING PROCEDURE                                     | VERSION |
| <b>DETECTION OF MEMBRANE PERMEABILITY BY<br/>7-AAD (LIVE DEAD MARKER)</b> | A       |

optimal use of this probe when it is combined with Fluorescein Isothiocyanate conjugated antibodies (FITC) and with R Phycoerythrin (PE) (3). Indeed, in contrast to Propidium Iodide (PI), which is another fluorescent probe used as a DNA marker, the 7-AAD/DNA complex has a reduced overlapping emission spectrum with FITC and PE.

### INTRA-LABORATORY REPRODUCIBILITY

On the same day and using the same cytometer, 12 independent measurements of the percentage of staining of a positive target (stabilized Stem-Trol Control Cells, 7-AAD positive) were carried out on a sample of whole fresh blood with positive target cells. The results obtained are summarized in the following table:

| Positive Target Stem-Trol | Number | Mean (%) | SD   | CV (%) |
|---------------------------|--------|----------|------|--------|
| 7-AAD <sup>+</sup>        | 12     | 3.63     | 0.55 | 15.2   |

### Linearity

To test the linearity of staining of this reagent, two cell types - Stem-Trol Control Cells (stabilized control cells, 7-AAD positive) and KG-1a cell line in culture (fresh line, ie 7-AAD negative) - were mixed in different proportions with a constant final number of cells, so that the Stem-Trol Control Cells / KG-1a ratio of the mixture ranged from 0 to 100%.

Aliquots were stained using the procedure described above and linear regression between the expected values and the observed values was calculated.

| Specificity | Linear regression | Linearity (R <sup>2</sup> ) |
|-------------|-------------------|-----------------------------|
| 7-AAD       | Y=1.0278X+4.1364  | 0.9704                      |

### LIMITATIONS

1. Flow cytometry may produce false results if the cytometer has not been aligned perfectly, if fluorescence leaks have not been correctly compensated for and if the regions have not been carefully positioned.
2. It is preferable to use a RBC lysis technique with a washing step as this reagent has not been optimized for "no wash" lysis techniques.
3. Accurate and reproducible results will be obtained as long as the procedures used are in accordance with the technical insert leaflet and compatible with good laboratory practices.
4. 7-AAD, the active substance in this reagent, has been optimized so as to offer the best specific/non-specific signal ratio. Therefore, it is important to adhere to the reagent volume/sample volume ratio in every test.
5. In the case of a hyperleucocytosis, dilute the blood in PBS so as to obtain a value of approximately  $5 \times 10^9$  leucocytes/L.
6. In certain disease states, such as severe renal failure or haemoglobinopathies, lysis of red cells may be slow, incomplete or even impossible. In this case, it is advisable to isolate mononucleated cells using a density gradient (Ficoll, for example) prior to staining.